

# Micros 60 Kontrollblatt für interne Qualitätskontrolle mit Minotrol 16 N

LOT: MX 415 N    Haltbar bis: 05.04.2019

*Nach dem Öffnen 14 Tage haltbar*

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.73	4.55 - 4.91							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	260	220 - 300							
MCV	79	75 - 83							
MCH	28.5	26.5 - 30.5							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.73	4.55 - 4.91							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	260	220 - 300							
MCV	79	75 - 83							
MCH	28.5	26.5 - 30.5							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.73	4.55 - 4.91							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	260	220 - 300							
MCV	79	75 - 83							
MCH	28.5	26.5 - 30.5							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									