

Micros 60 Kontrollblatt für interne Qualitätskontrolle mit Minotrol 16 N

LOT: MX 414 N Haltbar bis: 05.02.2019

Nach dem Öffnen 14 Tage haltbar

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	255	215 - 295							
MCV	80	76 - 84							
MCH	29.1	27.1 - 31.1							
MCHC	36.4	33.4 - 39.4							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	255	215 - 295							
MCV	80	76 - 84							
MCH	29.1	27.1 - 31.1							
MCHC	36.4	33.4 - 39.4							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.8	7.0 - 8.6							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	255	215 - 295							
MCV	80	76 - 84							
MCH	29.1	27.1 - 31.1							
MCHC	36.4	33.4 - 39.4							
Wöchentliche Reinigung									
Reagenz - Packwechsel									