

Micros 60 Kontrollblatt für interne Qualitätskontrolle mit Minotrol 16 N

LOT: MX 416 N Haltbar bis: 05.06.2019

Nach dem Öffnen 14 Tage haltbar

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	8.0	7.2 - 8.8							
RBC	4.68	4.50 - 4.86							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	265	225 - 305							
MCV	80	76 - 84							
MCH	28.8	26.8 - 30.8							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	8.0	7.2 - 8.8							
RBC	4.68	4.50 - 4.86							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	265	225 - 305							
MCV	80	76 - 84							
MCH	28.8	26.8 - 30.8							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	8.0	7.2 - 8.8							
RBC	4.68	4.50 - 4.86							
HGB	13.5	13.0 - 14.0							
HCT	37.4	34.9 - 39.9							
PLT	265	225 - 305							
MCV	80	76 - 84							
MCH	28.8	26.8 - 30.8							
MCHC	36.1	33.1 - 39.1							
Wöchentliche Reinigung									
Reagenz - Packwechsel									