

Micros EMI Kontrollblatt für interne Qualitätskontrolle mit Minotrol CRP

LOT: MC 409 Haltbar bis: 05.01.2019

Nach dem Öffnen 14 Tage haltbar

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.7 - 8.7							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	260	220 - 300							
MCV	80	76 - 84							
MCH	29.1	26.7 - 31.5							
MCHC	36.4	33.4 - 39.4							
CRP	25.0	17.0 - 33.0							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.7 - 8.7							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	260	220 - 300							
MCV	80	76 - 84							
MCH	29.1	26.7 - 31.5							
MCHC	36.4	33.4 - 39.4							
CRP	25.0	17.0 - 33.0							
Wöchentliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.7 - 8.7							
RBC	4.60	4.42 - 4.78							
HGB	13.4	12.9 - 13.9							
HCT	36.8	34.3 - 39.3							
PLT	260	220 - 300							
MCV	80	76 - 84							
MCH	29.1	26.7 - 31.5							
MCHC	36.4	33.4 - 39.4							
CRP	25.0	17.0 - 33.0							
Wöchentliche Reinigung									
Reagenz - Packwechsel									