

Celltac α 6400 & 6410 Kontrollblatt für interne Qualitätskontrolle mit MEK-3DN

LOT: B18YN Haltbar bis: 05.02.2019

Nach dem Öffnen 14 Tage haltbar

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.9 - 8.5							
RBC	4.68	4.48 - 4.88							
HGB	13.4	12.9 - 13.9							
HCT	39.8	36.8 - 42.8							
MCV	85.0	80.0 - 90.0							
MCH	28.6	26.2 - 31.0							
MCHC	33.7	30.7 - 36.7							
PLT	267	217 - 317							
14 Tägliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.9 - 8.5							
RBC	4.68	4.48 - 4.88							
HGB	13.4	12.9 - 13.9							
HCT	39.8	36.8 - 42.8							
MCV	85	80 - 90							
MCH	28.6	26.2 - 31.0							
MCHC	33.7	30.7 - 36.7							
PLT	267	217 - 317							
14 Tägliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.7	6.9 - 8.5							
RBC	4.68	4.48 - 4.88							
HGB	13.4	12.9 - 13.9							
HCT	39.8	36.8 - 42.8							
MCV	85	80 - 90							
MCH	28.6	26.2 - 31.0							
MCHC	33.7	30.7 - 36.7							
PLT	267	217 - 317							
14 Tägliche Reinigung									
Reagenz - Packwechsel									