

Celltac α 6318 Kontrollblatt für interne Qualitätskontrolle mit MEK-3DN

LOT: B18YN Haltbar bis: 05.02.2019

Nach dem Öffnen 14 Tage haltbar

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.9	7.1 - 8.7							
RBC	4.68	4.48 - 4.88							
HGB	13.3	12.8 - 13.8							
HCT	41.0	38.0 - 44.0							
MCV	87.5	82.5 - 92.5							
MCH	28.4	26.0 - 30.8							
MCHC	32.4	29.4 - 35.4							
PLT	264	214 - 314							
14 Tägliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.9	7.1 - 8.7							
RBC	4.68	4.48 - 4.88							
HGB	13.3	12.8 - 13.8							
HCT	41.0	38.0 - 44.0							
MCV	88	83 - 93							
MCH	28.4	26.0 - 30.8							
MCHC	32.4	29.4 - 35.4							
PLT	264	214 - 314							
14 Tägliche Reinigung									
Reagenz - Packwechsel									

Datum:	Mittel-	Sollbereich							
Visum:	wert	von - bis							
WBC	7.9	7.1 - 8.7							
RBC	4.68	4.48 - 4.88							
HGB	13.3	12.8 - 13.8							
HCT	41.0	38.0 - 44.0							
MCV	88	83 - 93							
MCH	28.4	26.0 - 30.8							
MCHC	32.4	29.4 - 35.4							
PLT	264	214 - 314							
14 Tägliche Reinigung									
Reagenz - Packwechsel									