



10745154196V34.0

Reflotron Precinorm U

REF 10745154 196

LOT 235709

2019-01

Components	Method	Value	Range	Units
Substrates				
BIL	Reflotron Bilirubin	1.50	1.23 - 1.77	mg/dL
		25.7	21.0 - 30.3	µmol/L
		15.0	12.3 - 17.7	mg/L
CREA	Reflotron Creatinine	2.02	1.72 - 2.32	mg/dL
		179	152 - 205	µmol/L
		20.2	17.2 - 23.2	mg/L
GLUC	Reflotron Glucose	113	96.1 - 130	mg/dL
		6.27	5.33 - 7.22	mmol/L
		1.13	0.96 - 1.30	g/L
UA	Reflotron Uric Acid	5.75	4.89 - 6.61	mg/dL
		342	291 - 393	µmol/L
		57.5	48.9 - 66.1	mg/L
UREA	Reflotron Urea ¹	45.3	37.1 - 53.5	mg/dL
		7.54	6.18 - 8.91	mmol/L
		0.45	0.37 - 0.54	g/L
Electrolytes				
K	Reflotron K ⁺	3.97	3.61 - 4.33	mmol/L
Lipides				
CHOL	Reflotron Cholesterol	190	163 - 217	mg/dL
		4.91	4.22 - 5.61	mmol/L
		1.90	1.63 - 2.17	g/L
TG	Reflotron Triglycerides	142	112 - 172	mg/dL
		1.62	1.28 - 1.97	mmol/L
		1.42	1.12 - 1.72	g/L

Components	Method		Value	Range	Units
Enzymes					
ALP ≥ 227520	Reflotron Alkal. Phosphatase	25; 30; 37 °C	102	80.6 - 123	U/L
			1.70	1.34 - 2.05	µkat/L
AMYL	Reflotron Amylase	25; 30; 37 °C	134	110 - 158	U/L
			2.23	1.83 - 2.63	µkat/L
P-AMYL	Reflotron Pancreatic Amylase	25; 30; 37 °C	95.0	77.9 - 112	U/L
			1.58	1.30 - 1.87	µkat/L



Reflotron Precinorm U

REF 10745154 196

LOT 235709

 2019-01

Components	Method		Value	Range	Units
Enzymes					
CK	Reflotron CK	25 °C	93.1	76.3 - 110	U/L
		30 °C	143	117 - 169	U/L
		37 °C	227	186 - 268	U/L
			3.78	3.10 - 4.47	µkat/L
AST	Reflotron GOT	25 °C	32.2	26.4 - 38.0	U/L
		30 °C	46.5	38.1 - 54.9	U/L
		37 °C	71.5	58.6 - 84.4	U/L
			1.19	0.98 - 1.41	µkat/L
ALT	Reflotron GPT	25 °C	29.0	23.8 - 34.2	U/L
		30 °C	41.6	34.1 - 49.0	U/L
		37 °C	54.7	44.9 - 64.5	U/L
			0.91	0.75 - 1.08	µkat/L
GGT	Reflotron GGT	25 °C	37.2	30.5 - 44.0	U/L
		30 °C	51.2	42.0 - 60.4	U/L
		37 °C	66.5	54.5 - 78.5	U/L
			1.11	0.91 - 1.31	µkat/L

1. Conversion factor: mg urea-N/dL = 0.467 x mg urea/dL / Umrechnungsfaktor: mg Harnstoff-N/dL = 0.467 x mg Harnstoff/dL / Facteur de conversion: mg d'azote uréique/dL = 0.467 x mg d'urée/dL ou g d'azote uréique/L = 0.467 x g d'urée/L / Factor de conversión: mg urea-N/dL = 0.467 x mg urea/dL / Fattore di conversione: mg di azoto ureico/dL = 0.467 x mg di urea/dL / omräkningsfaktor: mg Urea N/dL = 0.467 x mg Urea/dL / Omregningsfaktor: mg urea-N/dL = 0.467 x mg urea/dL / Omrekeningsfactor: mg ureum-N/dL = 0.467 x mg ureum/dL / Factor de conversão: mg ureia-N/dL = 0.467 x mg ureia/dL / Převodní faktor: mg N-močoviny/dL = 0.467 x močovina/dL / Препорачивачи: mg N-močoviny/dL = 0.467 x mg močoviny/dL / Współczynnik przeliczeniowy: azot mocznika mg/dL = 0.467 x mocznik mg/dL / Konverziós faktor: mg karbamid-N/dL = 0.467 x mg karbamid/dL / Συντελεστής μετατροπής: mg ουρίας-N/dL = 0.467 x mg ουρία/dL

Glossary / Glossar / Glossaire / Glosario / Glossario / Glossário / Ordliste / Ordlista / Ordliste / Verklarende woordenlijst / Glosář / Slovník / Słownik / Fogalomtár / Γλωσσάριο / Sözlükçe / Речник / Sónastik / Terminų žodynas / Vārdnīca / Глоссарий / 用語解説 :

Components / Bestandteile / Constituant / Componente / Componenti / Componente / Komponenter / Komponent / Komponent / Bestanddeel / Složky / Zložky / Składniki / Összetevők / Συστατικά / Bileşenler / Компоненти / Koostisosad / Komponentai / Sasitāvdaļas / Компоненты / 構成成分

Method / Methode / Méthode / Método / Metodo / Método / Methode / Metod / Methode / Method / Methode / Metoda / Metóda / Metoda / Eljárás / Μέθοδος / Yöntem / Метод / Meetod / Metodas / Metode / Метод / メソッド

Value / Wert / Valeur / Valor / Valore / Valor / Værdi / Värde / Verdi / Waarde / Hodnota / Hodnota / Wartość / Érték / Τιμή / Değer / Стойност / Väärtus / Verté / Vērtība / Значение / 平均值

Range / Bereich / Intervalle / Intervalo / Intervallo / Intervalo / Område / Intervall / Område / Bereik / Rozsah / Rozsah / Zakres / Tartomány / Εύρος / Aralık / Обхват / Vahemik / Intervalas / Diapazons / Диапазон / 範圍

Unit / Maßeinheit / Unité / Unjidad / Unità di misura / Unidade / Enhed / Enhet / Enhet / Meeteenheid / Jednotka / Jednotka / Jednostka / Mértékegység / Μονάδα / Birim / Единица / Ühik / Vienetas / Vienība / Единица / 單位

Substrates / Substrate / Substrats / Substratos / Substrati / Substratos / Substrater / Substrater / Substrater / Substraten / Substráty / Substráty / Substraty / Szubsztátok / Υποστρώματα / Substratlar / Субстрати / Substraadid / Substratai / Substrāti / Субстраты / 基質

Electrolytes / Elektrolyte / Electrolytes / Electrolitos / Elettroliti / Electrólitos / Elektrolytter / Elektrolyter / Elektrolytter / Electrolyt / Elektrolyty / Elektrolyty / Elektrolity / Elektrolitok / Ηλεκτρολύτες / Elektrolitler / Электролиты / Elektroliitid / Elektrolitai / Elektroliti / Электролиты / 電解質

Lipides / Lipide / Lipides / Lípidos / Lipidi / Lípidos / Lipider / Lipider / Lipider / Lipiden / Lipidy / Lipidy / Lipidy / Lipidek / Λιπίδια / Lipidler / Липиди / Lipiidid / Lipidai / Lipidi / Липиды / 脂質

Enzymes / Enzyme / Enzymes / Enzimas / Enzimi / Enzimas / Enzymer / Enzymer / Enzymer / Enzymen / Enzymy / Enzymy / Enzymy / Enzimek / Ένζυμα / Enzimler / Энзимы / Ensüümid / Fermentai / Enzimi / Ферменты / 酵素

© 2017, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

